

Referências bibliográficas

- [1] BRASIL. Lei n.º 9.478, de 6 de agosto de 1997. Dispõe sobre a política energética nacional, as atividades relativas ao monopólio do petróleo, institui o Conselho Nacional de Política Energética e a Agência Nacional do Petróleo e dá outras providências. **Diário Oficial da República Federativa do Brasil**, Brasília, DF, 7 ago. 1997. Disponível em: <www.anp.gov.br/conheca/lei.asp>. Acesso em 3 ago. 2009.
- [2] RTM. Portaria Conjunta n.º 1, de 19 de junho de 2000. Aprova o Regulamento Técnico de Medição de Petróleo e Gás Natural, que estabelece as condições e requisitos mínimos para os sistemas de medição de petróleo e gás natural, com vistas a garantir resultados acurados e completos. **Diário Oficial da República Federativa do Brasil**, Brasília, DF, 20 jun. 2000. Disponível em: <www.anp.gov.br/leg/inmetro.asp>. Acesso em 3 ago. 2009.
- [3] PETROLEO BRASILEIRO S.A. **Perfil**. Disponível em: <www2.petrobras.com.br/portugues/ads/ads_Petrobras.htm>. Acesso em 3 ago. 09.
- [4] AGÊNCIA NACIONAL DE PETROLEO, GÁS NATURAL E BIOCOMBUSTÍVEIS. **Glossário ANP**. Disponível em: <www.anp.gov.br/glossario/index.asp?strAlpha=M>. Acesso em 5 ago. 09.
- [5] _____. **Conheça a ANP**. Disponível em: <www.anp.gov.br/conheca/conheca.asp>. Acesso em 3 ago. 09.
- [6] Comunicação Particular.
- [7] AMERICAN GAS ASSOCIATION. **AGA Report No. 9: Measurement of Gas by Multipath Ultrasonic Meters**. Washington, DC, 2007.
- [8] _____. **AGA Report No. 10: Speed of Sound in Natural Gas and Other Related Hydrocarbon**: Washington, DC, 2003.
- [9] FOX, R. W.; McDONALD, A. T. **Introdução à Mecânica dos Fluidos**. Tradução de Alexandre Matos de Souza e Melo. 4 ed. Rio de Janeiro: Livros Técnicos e Científicos, 1998.
- [10] BRASIL, N. I. **Sistema Internacional de Unidades**. Rio de Janeiro: Interciência, 2002.
- [11] WYLEN, G. V.; SONNTAG, R.; BORGNAKKE, C. **Fundamentos da Termodinâmica Clássica**. Tradução de Euryale de Jesus Zerbini. 4 ed. São Paulo: Edgard Blücher, 1995.
- [12] AMERICAN GAS ASSOCIATION. **AGA Report No. 8: Compressibility Factors of Natural Gas and Other Related Hydrocarbon Gases**. Washington, DC, 1992.
- [13] NIST. **Fundamental Physical Constants**. Disponível em <http://physics.nist.gov/cgi-bin/cuu/Value?r|search_for=physchem_in!>. Acesso em 9 ago. 09.

- [14] ORLANDO, A. F. **Avaliação Metrológica de Medidores de Vazão de Petróleo e Gás Natural**. Rio de Janeiro, 2008
- [15] SANTIN, J. L. **Ultra-som: Técnica e Aplicação**. 2 ed. Curitiba: Artes Gráficas, 2003.
- [16] ORLANDO, A. F; VAL, L. G. **Calibração do Medidor de Vazão Krhone Altosonic V Offloading Marlim**. Relatório 1: Embasamento Teórico. Rio de Janeiro, 2006.
- [17] LANSING, J. et al. **Understanding and Using Ultrasonic Gas Flow Meters**. Houston (s.n.) [200-?].
- [18] SAVIDGE, S. L. **AGA10 Sound Speed Equations: Background, Thermodynamic Relations, Ideal Gas and Equation of State Methods, Uncertainty Analysis, and Calculation Flow Diagram for Natural Gas Measurement Applications**. In 5th International Symposium on Fluid Flow Measurement. Washington, D.C., 2002.
- [19] PETROBRAS TRANSPORTE S.A. **Atividades**. Disponível em: <<http://www.transpetro.com.br>>. Acesso em 9 ago. 09.
- [20] INMETRO. **Portaria INMETRO/DIMEL Nº 233**. Rio de Janeiro, 2005.
- [21] _____. **Vocabulário internacional de termos fundamentais e gerais de metrologia**. 4. ed. Rio de Janeiro, 2005.
- [22] ZANKER, K. J. **Transit Time Difference Ultrasonic Multi-path Gas Meters Diagnostics**. In: 2nd THE AMERICAS MEASUREMENT WORKSHOP, Houston, 2009.
- [23] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION. **ISO 17089: Measurement of Flow in Closed Conduits - Ultrasonic Meters for Gas – Part 1: Meters for Custody Transfer and Allocation Measurement**. Draft. Geneva, 2008.
- [24] AMERICAN NATIONAL STANDARDS INSTITUTE. **Overview**. Disponível em: <www.ansi.org>. Acesso em 9 ago. 09.
- [25] AMERICAN PETROLEUM INSTITUTE. **MPMS 21.1: Flow Measurement using Electronic Systems – Section 1: Electronic Gas Measurement**. Washington, DC, 1993.
- [26] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION. **ISO 12213: Calculation of compression factor - Part 3: Calculation using physical properties**. Geneva, 2006.
- [27] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION. **ISO VIML: International Vocabulary of Terms in Legal Metrology**. Geneva, 2000.
- [28] _____. **ISO 5168: Measurement of Fluid Flow – Procedures for the Evaluation of Uncertainties**. Geneva, 2005.
- [29] INMETRO. **NIT-DICLA-021: Expressão da Incerteza de Medição na Calibração**. Versão Brasileira da Publicação EA-4/02. Rio de Janeiro, 2003.
- [30] ORLANDO, A. F. **Uncertainty Analysis in Experimental Mechanics**. Rio de Janeiro, 200?
- [31] CEESI. **Flow Meter Calibration Facility (Colorado)**. Disponível em: <www.ceesi.com/docs/brochures/CEESI_Brochure.pdf>. Acesso em 9 ago. 09.

ANEXO A – Certificado de Calibração do medidor 08-060434

Folha 01/04



LABORATORY/OFFICE:
54043 County Rd. 37
Nunn, Colo. 80648
Phone: 970-897-2711
FAX: 970-897-2710

**COLORADO ENGINEERING
EXPERIMENT STATION INC.**

...the primary source for flow measurement solutions...



IOWA HIGH FLOW FACILITY
2365 240th St.
Garner, IA 50438
Phone: 641-923-3664
FAX: 641-923-3683

CERTIFICATE OF CALIBRATION
Traceable to the
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY

Customer: Daniel Measurement & Control **Item Calibrated:** 12" Ultrasonic Senior Sonic Flowmeter

Address: 11100 Brittmoore Park Drive **Model:** 3400
Houston, TX 77041

CEESI Data File(s): 08DAN-0429 **Serial Number:** 08-060434 (Electronics: 08-080700)

Test Date: 20 March 2008 **Job Number:** CE06692 **Order:** 195494-000 OP

The uncertainty in flowrate indicated by CEESI standards is estimated to be +/- 0.23% of reading to 95% confidence.

The calibration identified by the above CEESI DATA FILE was performed in accordance with the American Gas Association Transmission Committee Report Number 9 using calibration standards that are traceable to the National Institute of Standards and Technology. All natural gas properties were calculated in compliance with American Gas Association Report Number 8.

This Calibration is: As Found

As Left

Measurement Assurance

On behalf of Colorado Engineering
Experiment Station, Inc.

NOTE: This Certificate and accompanying documentation shall not be reproduced, except in full, without the written consent of Colorado Engineering Experiment Station Inc.

Folha 02/04



LABORATORY/OFFICE:
54043 County Rd. 37
Nunn, Colo. 80648
Phone: 970-897-2711
FAX: 970-897-2710

COLORADO ENGINEERING EXPERIMENT STATION INC.

...the primary source for flow measurement solutions...



IOWA HIGH FLOW FACILITY
2365 240th St.
Garner, IA 50438
Phone: 641-923-3664
FAX: 641-923-3693

Calibration of a 12" Ultrasonic Senior Sonic Flowmeter

Model: 3400

Serial Number: 08-060434 (Electronics: 08-080700)

Bore Diameter: 11.3735 inches

Test Location: Ventura Iowa

Test Medium: Natural Gas

Test Date: 20 March 2008

CEESI Data File: 08DAN-0429

Nominal Flowing Pressure: 1080 psia

Nominal Flowing Temperature: 68 °F

Gas Composition During Test:

<u>Component</u>	<u>Beginning Normalized Mole Percent</u>	<u>Ending Normalized Mole Percent</u>
Methane	95.4505	95.4496
Nitrogen	1.4651	1.4582
Carbon Dioxide	0.7219	0.756
Ethane	2.0639	2.044
Propane	0.1024	0.1072
Hydrogen	0.1574	0.1456
Iso-Butane	0.0032	0.003
Normal-Butane	0.004	0.0041
Iso-Pentane	0.0	0.0
Normal-Pentane	0.0	0.0
C6+	0.0	0.0
Helium	0.0317	0.0323
Heating Value (BTU/ft ³)	1008.2771	1007.9973
Total Un-Normalized	100.0228	100.1875

Folha 03/04



LABORATORY/OFFICE:
54043 County Rd. 37
Nunn, Colo. 80648
Phone: 970-897-2711
FAX: 970-897-2710

COLORADO ENGINEERING EXPERIMENT STATION INC.

...the primary source for flow measurement solutions...



IOWA HIGH FLOW FACILITY
2365 240th St.
Garner, IA 50438
Phone: 641-923-3664
FAX: 641-923-3693

Calibration Results

The calibration results are summarized in the table below. The column labeled 'Calibration Factor' is the value by which the volume flowrate is multiplied to achieve the corrected value. The column labeled 'Random Effects' is the uncertainty (at 95% confidence) in average Calibration Factor due to all random effects (repeatability, reproducibility and hysteresis) observed during the calibration process.

Flowrate [%FS]	Flowrate [acfh]	Velocity [ft/sec]	Calibration Factor	Random Effects[%]
100.51	255276	100.5	1.0050	0.033
70.93	180143	70.9	1.0041	0.012
40.79	103613	40.8	1.0034	0.013
20.49	52041	20.5	1.0026	0.046
10.56	26822	10.6	1.0025	0.096
2.56	6510	2.6	1.0029	0.068

The uncertainty due to systematic effects present in the calibration facility is estimated to be $\pm 0.23\%$ at a 95% level of confidence.

Upon Completion of the calibration the coefficients listed below were input into the meter and then the meter was put into a "Read Only" mode.

Register	Name	Value	Register	Name	Value
328	FwdFlwRt1	254006	340	FwdMtrFctr1	1.005
329	FwdFlwRt2	179407	341	FwdMtrFctr2	1.0041
330	FwdFlwRt3	103262	342	FwdMtrFctr3	1.0034
331	FwdFlwRt4	51906	343	FwdMtrFctr4	1.0026
332	FwdFlwRt5	26756	344	FwdMtrFctr5	1.0025
333	FwdFlwRt6	6491	345	FwdMtrFctr6	1.0029
334	FwdFlwRt7	0	346	FwdMtrFctr7	1
335	FwdFlwRt8	0	347	FwdMtrFctr8	1
336	FwdFlwRt9	0	348	FwdMtrFctr9	1
337	FwdFlwRt10	0	349	FwdMtrFctr10	1
338	FwdFlwRt11	0	350	FwdMtrFctr11	1
339	FwdFlwRt12	0	351	FwdMtrFctr12	1

Folha 04/04.



LABORATORY/OFFICE:
54043 County Rd. 37
Nunn, Colo. 80648
Phone: 970-897-2711
FAX: 970-897-2710

COLORADO ENGINEERING EXPERIMENT STATION INC.


...the primary source for flow measurement solutions...



IOWA HIGH FLOW FACILITY
2365 240th St.
Garner, IA 50438
Phone: 641-923-3664
FAX: 641-923-3693

Calibration Results


Daniel Measurement & Control
12" Ultrasonic Senior Sonic Flowmeter
S/N: 08-060434 (Electronics: 08-080700)
CEESI Iowa Calibration Curve



IBR TRANSPETRO		CERTIFICADO DE CALIBRAÇÃO Nº: 12/05/2009						
Transmissor de Temperatura		<input checked="" type="checkbox"/> Calibração em Malha Fechada						
Ponto de Medição: ETC Cacimbas		Data da calibração: 12-mai-09						
IDENTIFICAÇÃO DO INSTRUMENTO								
Tag: TIT-4150.4422A		Modelo: 3144PD1A1E1B4M504						
Fabricante: ROSEMOUNT		Série: 0587385						
CARACTERÍSTICAS DO INSTRUMENTO								
Faixa operacional:	a 29,48	Classe de Exatidão:	0,0010					
Faixa de Calibração:	a 100,00	Menor Divisão da Escala:	0,0010					
DADOS COMPLEMENTARES								
Procedimento de Calibração: PE-3N3-00657		Incerteza padronizada resultante da calibração do ohmímetro em % do valor lido						
Condições Ambientais: 27°C		u(Ometro)						
PADRÃO(ÕES) UTILIZADO(S)								
Identificação	Tag	Erro Sistemático (Tendência)	Incerteza Expandida	k (certificado)	Incerteza Padrão	Valor de uma divisão	Nº Certificado	Laboratório
Calib. Digital de temp. Presys-T25N-2-BP06			0,0300	2	0,015	0,0100	R1775.10.07	Pyrne Lab
Calib. Digital de temp. Presys-T25N-2-BP06			0,0800	2	0,040	0,0100	R1776.10.07	Pyrne Lab
				2	0,000			
				2	0,000			
RELATÓRIOS DE ENSAIO								
AVALIAÇÃO INICIAL								
1ª leitura - ascendente	2ª leitura - descendente	1ª leitura - ascendente	2ª leitura - descendente	AVALIAÇÃO FINAL				
padrão	instrumento	padrão	instrumento	padrão	instrumento	Histerese	Desvio Padrão	
0,094	0,031	0,000	0,031	-	-	0,031	0,022	
29,953	30,000	30,078	30,047	-	-	0,031	0,022	
49,922	50,000	49,922	49,953	-	-	0,031	0,022	
79,906	80,000	79,938	80,000	-	-	0,062	0,044	
99,969	100,000	100,000	100,000	-	-	0,000	0,000	
AVALIAÇÃO DE INCERTEZA				<p>PREENCHER O CAMPO: UNIDADE DO PADRÃO E UNIDADE DO INSTRUMENTO, CORRETAMENTE.</p>				
Incerteza Combinada do(s) Padrão(ões)		0,043		Incerteza Combinada Total		0,103		Erro Sistemático
Resolução(ões) do Padrão(ões)		0,004		Cálculo do Verif		85,23		Remanescente
Repetitividade do Instrumento		0,044		k Corrigido		1,99		0,016
Resolução do Instrumento		0,000		Incerteza Expandida		0,20		0,063
Incerteza da Histerese		0,018		Critério de aceitação		1,00		0,063
Incerteza do Ajuste da Curva de Calibração		0,050		Avaliação Inicial		APROVADO		0,031
				Avaliação Final		APROVADO		0,000
				INSTRUMENTO AJUSTADO?				
				NAO				
REALIZADO POR: Fabiano/Francisco Emílio				SUPERVISOR: Evaristo				
OBSERVAÇÃO:								

ANEXO C – Certificado de verificação sem escoamento.

Folha 01/02

Specification Verification For Daniel Mark III Ultrasonic Meter**Customer And Unit Information:**

Customer: Emerson Brazil
 Sales Order: 283886
 Electronics Housing Serial No: 08-080700
 Meter Housing Serial No: 08-060434
 Modbus ID: 32
 Meter Size: 12" 600
 CPUBdSwVer: 1 60-Gas_Release_Prod 2007/11/08
 OSVer: 2.4.21-USM-1 2007/7/13
 FileSysVer: 2 02 2007/03/19
 DatabaseConfigVersion: 782
 Device Number: 3400

Calibration Equipment Data:

Pressure Transducer Serial No: 0382554
 Pressure Transducer Calib date: 11/29/07
 Temperature Transducer Serial No: 1098493
 Temperature Transducer Calib date: 11/29/07
 RTD Probe Serial No: 147291
 RTD Probe Calib date: 11/29/07
 DUI Software Version: 2.23
 Daniel CUI Software Version: 4.0

Meter Dimensions:

XA (inch) : 3.8553
 XB (inch) : 6.2468
 XC (inch) : 6.2453
 XD (inch) : 3.8615
 LA (inch) : 8.9598
 LB (inch) : 13.0778
 LC (inch) : 13.0811
 LD (inch) : 8.9585
 CMM Bore Dia (inch) : 11.3735

Waveforms:

Are Waveforms A, B, C & D clean with positive 1st peaks: Yes


Date: 3/18/2008
 Time: 12:56 PM

08-060434.doc

1 of 2

Folha 02/02

Specification Verification For Daniel Mark III Ultrasonic Meter



Meter Delays:

AvgDlyA (us) : 19.953
 AvgDlyB (us) : 20.271
 AvgDlyC (us) : 20.542
 AvgDlyD (us) : 20.090
 DltDlyA (us) : 0.001
 DltDlyB (us) : -0.030
 DltDlyC (us) : -0.001
 DltDlyD (us) : -0.001

Readings At Zero Flow:

Temperature(deg F): 75.5
 Pressure (psi): 194.94
 Atmospheric Pressure (psi): 14.70
 Speed of Sound (ft/sec): 1160.70
 Recorded Flow velocity (ft/sec): -0.007
 Recorded Flow Rate (ft³/hr): -18.96

Meterology Data Verification:

Bore Diameter check: PASSED
 X Dimension check: PASSED
 L Dimension check: PASSED
 AvgDly check: PASSED
 DltDly check: PASSED

Flow Related Data Verification:

FlowVels are within tolerance limit ($S_r = 0.03$, $I_r = 0.05$)ft/sec: PASSED
 Avgflow is within tolerance limit ($S_r = 0.014$, $I_r = 0.030$) ft/sec: PASSED
 SndVels are within tolerance limit of 1.0 ft of each other: PASSED
 Gains are within tolerance limit of 15->70 dB: PASSED
 NEs are within tolerance limit of 250: PASSED
 SDevTms are within tolerance limit of 100 nano seconds: PASSED
 SDevDltTms are within tolerance limit of 100 nano seconds: PASSED
 DltTms are within tolerance limit of 0.050 micro seconds: PASSED
 Performance of all working chords is at 100 %: PASSED

Tested by : Carl Beldh
 Approved by: _____

Date : 3-18-08
 Date : _____

Date: 3/18/2008
 Time: 12:56 PM

08-060434 doc

2 of 2

ANEXO D – Projeto da estação de medição para transferência de custódia onde os medidores utilizados neste dissertação encontram-se instalados.

